

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Amplification using the Power Plex Fusion System		
Status:Published		Document ID: 5980
DATE EFFECTIVE 08/27/2019	APPROVED BY Nuclear DNA Technical Leader	PAGE 1 OF 4

PowerPlex® Fusion Sample Preparation for Amplification

1 Procedure

Fusion Sample Input Amount
Optimal – 525pg*
Minimum – 37.5pg

*The option for amplification with a greater input amount is available if determined appropriate for the sample by the analyst.

- 1.1 Retrieve the following reagents from the associated refrigerator and/or freezer.

PowerPlex Fusion® 5X Primer Pair Mix
PowerPlex Fusion® 5X Master Mix
Water, Amplification Grade
2800M Control DNA, .250ng/µl

- 1.2 Retrieve sample(s) needed for amplification from associated refrigerator and/or freezer.
- 1.3 Prepare dilutions **using Promega Amplification Grade Water**, for each sample, if necessary, according to Table 1. Vortex and centrifuge samples prior to aliquoting for dilution.

TABLE 1: Dilutions

Dilution	Amount of DNA Template (uL)	Amount of Promega® Water (uL)
0.25	3 or (2)	9 or (6)
0.2	2	8
0.1	2	18
0.05	2	38
0.04	4 or (2)	96 or (48)
0.02	2 or (1)	98 or (49)
0.01	2	198
0.008	4 or (2)	496 or (248)

- 1.4 Centrifuge reagent tubes briefly to bring contents to the bottom and then vortex for 15 seconds before use. Do NOT re-centrifuge the Master Mix or Primer Pair Mix as this may cause the reagents to be concentrated at the bottom of the tube.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Amplification using the Power Plex Fusion System		
Status:Published		Document ID: 5980
DATE EFFECTIVE 08/27/2019	APPROVED BY Nuclear DNA Technical Leader	PAGE 2 OF 4

- 1.5 Consult the Reagents tab in LIMS for the exact amount of PowerPlex Fusion® 5X Primer Pair Mix and PowerPlex Fusion® 5X Master Mix to add.

Reagent	Per reaction
5X Primer Pair Mix	2.5 µL
5X Master Mix	2.5 µL
Mastermix total:	5 µL
DNA	7.5 µL

- 1.6 Vortex prepared Master Mix and samples created in Step 3. After vortexing, **briefly centrifuge** master mix and samples.
- 1.7 Add **5 µL** of the prepared master mix to each tube that will be utilized, changing pipette tips and remixing master mix as needed.
- 1.8 **Witness Step.** Have another analyst witness the sample set-up.
- 1.8.1** For the input samples, confirm the tube label and sample ID for each sample. For the output samples, the entire amp tube label must be read for each sample.
- 1.9 Positive Control – total input amount of 500pg.
- 1.9.1 Aliquot positive control according to amplification sheet
- 1.10 Amplification Negative
- 1.10.1 7.5 uL of Water, Amplification Grade
- 1.11 Samples
- 1.11.1 Aliquot samples according to amplification sheet
- 1.12 Ensure that all caps are properly closed prior to sending the samples to the post-amplification laboratory.
- 1.13 Spin down samples at 1000 RPM for one minute.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Amplification using the Power Plex Fusion System		
Status:Published		Document ID: 5980
DATE EFFECTIVE 08/27/2019	APPROVED BY Nuclear DNA Technical Leader	PAGE 3 OF 4

2 PowerPlex® FusionPCR Conditions for the Applied Biosystems GeneAmp PCR System 9700

- 2.1 Turn on the ABI 9700 Thermal Cycler.
- 2.2 Choose the following program in order to amplify these samples:

PowerPlex® Fusion
user: casework
file: PPFusion-29

- 2.3 PowerPlex® Fusion PCR Conditions for the Applied Biosystems GeneAmp PCR System 9700

9700	The PowerPlex® Fusion file is as follows:
PowerPlex® Fusion	Soak at 96°C for 1 minutes
user: casework	29 Cycles : Denature at 94°C for 10 seconds
file: ppfusion-29	: Anneal at 59°C for 60 seconds
	: Extend at 72°C for 30 seconds
	10 minute incubation at 60°C.
	Storage soak indefinitely at 4°C

- 2.4 Record instrument in LIMS
- 2.5 The run will start when the heated cover reaches temperature. The screen will then display a flow chart of the run conditions. A flashing line indicates the step being performed, hold time is counted down. Cycle number is indicated at the top of the screen, counting up.
- 2.6 Upon completion of the amplification:
 - 2.6.1 Remove samples and press the STOP button repeatedly until the “End of Run” screen is displayed.
 - 2.6.2 Select the EXIT option (F5).
 - 2.6.3 Wipe any condensation from the heat block with a lint free wipe and pull the lid closed to prevent dust from collecting on the heat block.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Amplification using the Power Plex Fusion System		
Status:Published		Document ID: 5980
DATE EFFECTIVE 08/27/2019	APPROVED BY Nuclear DNA Technical Leader	PAGE 4 OF 4

- 2.6.4 Turn the instrument off.
- 2.7 Place the microtube rack used to set-up the samples for PCR in the container of 10% bleach container in the Post-Amp area.
- 2.8 After completion of the thermal cycling protocol, store amplified product at 4°C and proceed with fragment analysis for exemplars.
- 2.9 Complete the LIMS test batch
 - 2.9.1 Fill out the performed by tab for the Amplification Run Review.
 - 2.9.2 Select all output samples and click Review to perform the test batch approval.
- 2.10 Schedule the samples to the appropriate STR test batch and create the test batch.

ARCHIVED