

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

PowerPlex Fusion - Capillary Electrophoresis		
Status: Published		Document ID: 5981
DATE EFFECTIVE 06/25/2018	APPROVED BY Nuclear DNA Technical Leader	PAGE 1 OF 5

PowerPlex Fusion – Capillary Electrophoresis

1 Procedure

- 1.1 Retrieve amplified samples from the thermal cycler or refrigerator.
- 1.2 Prepare thermal cyclers for snap dechill step.
 - 1.2.1 Set one thermal cycler to 95C (heat program) and one thermal cycler to 4C (chill program).
 - 1.2.1.1 Thermal cyclers should be turned off after each use.
- 1.3 Spin down samples at 1000 RPM for one minute.
- 1.4 Retrieve the following reagents from the associated refrigerator and/or freezer.

PowerPlex Fusion® WEN ILS 500
PowerPlex Fusion® Allelic Ladder
HiDi Formamide*

- 1.4.1 HiDi Formamide must not be re-frozen.
- 1.5 Record lot numbers of reagents.
- 1.6 Prepare one mastermix for all samples, negative and positive controls, and allelic ladders as specified below:

(9.5 µL of HiDi + 0.5 µL of WEN ILS500 per sample)

# Samples + 2	HiDi Form	ILS500
16	171 uL	9 uL
32	323 uL	17 uL
48	475 uL	25 uL
64	627 uL	33 uL
80	779 uL	41 uL
96	931 uL	49 uL

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

© NYC OFFICE OF CHIEF MEDICAL EXAMINER

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

PowerPlex Fusion - Capillary Electrophoresis		
Status:Published		Document ID: 5981
DATE EFFECTIVE 06/25/2018	APPROVED BY Nuclear DNA Technical Leader	PAGE 2 OF 5

- 1.7 Obtain a reaction plate and label the side with the run name.
- 1.8 Aliquot **10 μ L** of **mastermix** to each well.
 - 1.8.1 If an injection has less than 16 samples, add at least 10 uL of either dH₂O, formamide, HiDi, buffer, or mastermix to all unused wells within that injection.
- 1.9 **Witness step.** Have another analyst verify the tube setup by comparing the tube labels and positions indicated on the Load Plate Screen in the LIMS system with the tube labels and positions of the amp tubes.
 - 1.9.1 For samples being transferred from a Fusion Direct amplification plate:
 - 1.9.1.1 The witness must verify that each individual sample well is in the same order by comparing the amplification test batch Load Plate screen against the STR test batch Load Plate screen. This must be verified by hovering the cursor over the wells and corresponding samples on the plate images, not by using the load plate sample list.
 - 1.9.1.2 The analyst must state the test batch ID on the amplification plate and amplification time; the witness must verify this information in the amplification test batch in LIMS.
 - 1.9.1.3 Controls- verify that the positive control date and time from the amplification test batch is the same as the positive control date and time in the STR test batch.
 - 1.9.1.4 Sample wells- analyst setting up the plate must state which wells hold amplified product and are being transferred from the amplification plate. Analyst must also state the wells being loaded in the STR plate. Witness must verify that the wells are correct in the Load Plate screens of the amplification test batch and the STR test batch.
- 1.10 Aliquot **1 μ L** of allelic ladder, positive control, negative control, and sample into their appropriate well.
 - 1.10.1 Sample order is as follows: A1, B1, C1... A2, B2, C2, etc.
 - 1.10.2 Samples may be loaded from a PowerPlex Fusion Direct amplification plate into the CE plate by using a multichannel pipette. **1 μ l of allelic ladder must be loaded separately into the appropriate wells.**
- 1.11 Once all of the samples have been added to the plate, place a new 96-well septa over the reaction plate and firmly press the septa into place.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

PowerPlex Fusion - Capillary Electrophoresis		
Status: Published		Document ID: 5981
DATE EFFECTIVE 06/25/2018	APPROVED BY Nuclear DNA Technical Leader	PAGE 3 OF 5

- 1.12 Spin the plate in the centrifuge at 1000 RPM for one minute.
- 1.13 Denature plate for **3 minutes** with the thermal cycler set to 95°C.
- 1.14 Chill plate for **3 minutes** with the thermal cycler set to 4°C.
- 1.15 While plates are denature/chilling set up 3130 for run.
 - 1.15.1 Turn on oven and set for 60°
 - 1.15.1.1 Manual Control → Send Defined Command For: click on Oven.
 - 1.15.1.2 Command Name click on Turn On/Off oven → Send Command
 - 1.15.1.3 Command Name click on Set Oven Temperature → 60 → Send Command
 - 1.15.2 Check if instrument needs a POP4 change. POP4 must be changed after seven days. POP4 change is needed if there is not enough to run the plate. A full piston chamber is ~600µL. A 6 injection plate requires ~450µL.
 - 1.15.3 **DO NOT OPEN INSTRUMENT DOORS IF AUTOSAMPLER IS IN MOTION.**
 - 1.15.3.1 Remove a new bottle of POP4 from the refrigerator.
 - 1.15.3.2 Select **Wizards** → **Water Wash Wizard** and follow the wizard.
 - 1.15.3.3 Note: when asked to turn the array knob, only turn the knob if bubbles are present. If the knob is turned, only turn it a quarter turn and ensure to re-tighten the knob once complete. If the capillary appears to twist, contact QA.
 - 1.15.3.4 When the **Fill Array** step has completed, remove the anode buffer jar, empty, and fill with 1X Buffer (~15 mL).
 - 1.15.3.5 Close instrument doors and wait for the steady green light.
 - 1.15.3.6 Click **Finish**
 - 1.15.4 Change buffer on instrument. (Buffer only needs to be changed once daily). **DO NOT OPEN INSTRUMENT DOORS IF AUTOSAMPLER IS IN MOTION.**
 - 1.15.4.1 Remove reservoirs and anode buffer jar from instrument.
 - 1.15.4.2 Rinse all containers using distilled water and wipe dry with a lint free wipe.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

PowerPlex Fusion - Capillary Electrophoresis		
Status:Published		Document ID: 5981
DATE EFFECTIVE 06/25/2018	APPROVED BY Nuclear DNA Technical Leader	PAGE 4 OF 5

- 1.15.4.3 Make 1X buffer (45 ml Ultra Pure water, 5 ml 10X buffer) in a 50 mL conical tube
- 1.15.4.4 Fill the buffer jar and anode buffer jar with 1X Buffer to their respective lines.
- 1.15.4.5 Fill the waste and water reservoirs with Ultra Pure water to their respective lines.
- 1.16 Import text file
 - 1.16.1 Copy the LIMS plate record from the LIMS fileshare to the Plate Records folder on the instrument computer.
 - 1.16.2 Plate Manager → Import → D:\AppliedBiosystems\Plate Records → Plate Record for Current Run
- 1.17 Set G5 spectral to most recent PPFusion spectral.
 - 1.17.1 Spectral viewer → Dye Set drop down select G5 → List of Calibrations for Dye Set: G5 select PPFusionG5 spectral with the most recent date → click Set
 - 1.17.2 This spectral must be changed to Yfiler of Minifiler G5 spectral before running a Yfiler or Minifiler plate on this instrument. Due to the need to change spectrals between Fusion and Yfiler/Minifiler, Yfiler or Minifiler samples CANNOT be run on the same plate, or as the second plate on an instrument running a Fusion plate.
- 1.18 Spin down plate at 1000RPM for 1 minute.
- 1.19 Linking plate to instrument
 - 1.19.1 Run Scheduler → Plate View
 - 1.19.2 Push the tray button on the bottom left of the machine and wait for the autosampler to move forward and stop at the forward position.
 - 1.19.3 Open the doors and place the tray onto the autosampler in the correct tray position, A or B. There is only one orientation for the plate. (The notched end faces away from the user.
 - 1.19.4 Ensure the plate assembly fits flat in the autosampler. The Plate View window will change from gray to yellow.
 - 1.19.5 Type the exact plate name in the Plate ID window and click “Search.” Or, click the “Find All” button and select the desired plate record.

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

PowerPlex Fusion - Capillary Electrophoresis		
Status:Published		Document ID: 5981
DATE EFFECTIVE 06/25/2018	APPROVED BY Nuclear DNA Technical Leader	PAGE 5 OF 5

- 1.19.6 Click on plate record and then corresponding plate position in the Plate View screen. This will cause the yellow screen to turn green.
- 1.20 Check Run View to ensure correct number of samples and injections are being run.
- 1.20.1 Run Scheduler → Run View
- 1.20.2 Click on the run file to see the Plate Map or grid diagram.
- 1.21 Ensure instrument is ready to go by checking reservoirs are filled properly, plate is linked correctly, and no bubbles are present in the polymer block.
- 1.21.1 If bubbles are present proceed to **Bubble Removal Wizard**. (Wizards→Bubble Removal Wizard)
- 1.22 Start run by pressing green **Run** button (play button).
Monitor instrument until event log shows the instrument as begun Pre-Run.

	U
Oven Temp	60°C
Pre-Run Voltage	15.0 kV
Pre-Run Time	180 sec
Injection Voltage	3 kV
Injection Time	5 sec
Run Voltage	13 kV
Run Time	2000 sec

- 1.23 Enter a LIMS usage log for current run, recording lot numbers of buffer and POP4 in the usage log if they were changed during for run.
- 1.24 Amplification tubes should be stored at 4°C-8°C, filed by instrument and run name. Fusion Direct amplification plates should be discarded.