

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

ReRun Codes		
Status:Published		Document ID: 1123
DATE EFFECTIVE 05/02/2017	APPROVED BY Nuclear DNA Technical Leader	PAGE 1 OF 1

SAMPLE COMMENTS INDEX

Code Groups	Parameters for Rerun	insert sample status code (if applicable)	insert system code	insert rerun code (if applicable)	insert realiquot code (if applicable)
A) Sample Status Code	All peak labels removed.	**			
	Peak(s) within basepair range affected by overblown peak(s). Labels removed.*	*			
	Sample chart shows OL/OB allele	^			
	No or poor size standard	#			
B) System Code	Identifiler®		I		
	Minifiler®		F		
	Yfiler®		M		
	PowerPlex Fusion®		U		
	Do Not Rerun		N/A		
C) Rerun Code	Normal (HCN)			no code	
	High (HCN)*			R	
	1 kV 22 s (LCN)			L	
	3 kV 20 s (LCN)			N	
	6 kV 30 s (LCN)			H	
D) Realiquot Code	1/5 dilution				D.2
	1/10 dilution				D.1
	1/20 dilution				D.05
	1/100 dilution				D.01
	Re-aliquot 1 ul				1ul
	Re-aliquot 2 ul				2ul

If multiple rerun codes apply, place a comma between each in the sample comments (i.e. an YF system sample's chart shows OL/OB alleles and some unlabeled peaks just below 75 RFU threshold that could be part of the contributor's profile, to schedule the sample for rerun normal and high, enter “^M,MR” in the sample comments)

*Not for use with Fusion®

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.