

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

ReRun Codes		
<small>Status:Published</small>		<small>Document ID: 1123</small>
DATE EFFECTIVE 12/20/2023	APPROVED BY Nuclear DNA Technical Leader	PAGE 1 OF 1

SAMPLE COMMENTS INDEX

Code Groups	Parameters for Rerun	insert sample status code (if applicable)	insert system code	insert rerun code (if applicable)	insert realiquot code (if applicable)
A) Sample Status Code	All peak labels removed.	**			
	Sample chart shows OL/OB allele	^			
	No or poor size standard	#			
	Excessive elevated baseline	^^			
	Excessive spikes	##			
	Migration shift	+			
B) System Code	PowerPlex Fusion®		U		
	PowerPlex Y23®		P		
	Do Not Rerun		no code		
C) Realiquot Code	1/5 dilution				D.2
	1/10 dilution				D.1
	1/20 dilution				D.05
	1/100 dilution				D.01
	No dilution				no code

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.